



# A&ZMED

Personal Protective Equipment

## OLI-2022 MASK

*Non-woven, triple layer, disposable mask.*



Fabric surface blocks liquids within a special technology.



Our products prevent perspiration by their breathability characteristic, and provide a healthy and comfortable use.



Prevent the expansion and growth of bacterias.



**STANDARDS: EN 14683:2019+AC:2019**



ISO-9001:2015



ISO-13485:2016




**LAYER -1**

PRODUCT	%100 Polypropylene Spunbond	
COLOR	BLUE / WHITE	
WEIGHT	20	
MADE TEST	AVARAGE VALUE +/- %10	METHOD
TENSILE STRENGTH N / 5 cm (MD)	39,44	NWSP 110.1 R0(15)
TENSILE STRENGTH N / 5 cm (CD)	27,10	NWSP 110.1 R0(15)
ELONGATION % (MD)	72,00	NWSP 110.1 R0(15)
ELONGATION % (CD)	81,00	NWSP 110.1 R0(15)
WEIGHT (gr/m2)	20,0	NWSP 130.1 R0(15)
STRIKE THROUGH TIME (s)	0,00	NWSP 070.7 R0(15)
REWET (g)	0,00	NWSP 070.9 R1(15)

**LAYER -3**

PRODUCT	%100 Polypropylene Spunbond	
COLOR	BLUE / WHITE	
WEIGHT	30	
MADE TEST	AVARAGE VALUE +/- %10	METHOD
TENSILE STRENGTH N / 5 cm (MD)	66,30	NWSP 110.1 R0(15)
TENSILE STRENGTH N / 5 cm (CD)	41,20	NWSP 110.1 R0(15)
ELONGATION % (MD)	81,20	NWSP 110.1 R0(15)
ELONGATION % (CD)	88,50	NWSP 110.1 R0(15)
WEIGHT (gr/m2)	30,0	NWSP 130.1 R0(15)
STRIKE THROUGH TIME (s)	0,00	NWSP 070.7 R0(15)
REWET (g)	0,00	NWSP 070.9 R1(15)



**LAYER -2**

PRODUCT	%100 Polypropylene Meltblown	
COLOR	WHITE	
WEIGHT	25	
MADE TEST	AVARAGE VALUE +/- %10	METHOD
TENSILE STRENGTH N / 5 cm (MD)	13,80	NWSP 110.1 R0(15)
TENSILE STRENGTH N / 5 cm (CD)	12,40	NWSP 110.1 R0(15)
ELONGATION % (MD)	42,00	NWSP 110.1 R0(15)
ELONGATION % (CD)	53,60	NWSP 110.1 R0(15)
WEIGHT (gr/m2)	25,0	NWSP 130.1 R0(15)
HYDROSTATIC HEAD (mmH <sub>2</sub> O)	412,00	NWSP 080.6 R0(15)
BACTERIAL FILTRATION EFFICIENCY, BFE (%)	≥ 95	EN 14683:2019





# A&ZMED

Personal Protective Equipment

## OLI-2022 MASK

### (Instructions)

- 1) Open the package, take out the mask, hold the two ends of the mask with the bridge of the nose facing upwards;
- 2) The mask covers the mouth, nose, and jaw, and the earband is placed behind the ear;
- 3) Press the bridge strip to fit the bridge of the nose;
- 4) Adjust the position of the mask appropriately to make it comfortable to wear;
- 5) Wear correctly according to the front and back sides. Do not touch the inside of the mask with your hands.

### (Precautions)

- 1) Please read the instructions carefully to avoid touching the inner layer of the mask with your hands;
- 2) Non-sterile masks should not be used in areas with strict microbiological index control;
- 3) Please check the packing before use, if the packing is damaged, it is forbidden to use;
- 4) This product is disposable and cannot be used repeatedly;
- 5) If liquid splashing, mask damage, dampness, or significant increase in breathing resistance occur during wearing, please replace it in time;
- 6) After using the product, it should be disposed of in accordance with the requirements of the hospital or environmental protection department.

## How to Wear a Mask



## How to Remove Masks



Made in Turkey

İBİŞLER TEKSTİL A.Ş.

Orhangazi Mah. Tunç Cad. No:5 34538 Esenyurt / İstanbul - TURKEY

+90 212 602 04 05 www.ibisler.com info@ibisler.com



8682742370706



## ATTESTATION OF CONFORMITY

Certificate Nr: MDD - 098

In conformance to the European Economic Commission 93/42/EEC Medical Devices Directive on harmonisation of laws, regulations and administrative documentation of Member States on Medical Devices and European Economic Commission directive 93/68/EEC amending Medical Devices Directive dated 22 July 1993,

the products manufactured by

**İBİŞLER TEKSTİL SANAYİ VE DIŞ TİCARET ANONİM ŞİRKETİ**

at the following address

Orhangazi Mahallesi Tunç Caddesi No:5 34358 Esenyurt / İSTANBUL

**EN 14683:2019+AC:2019 Medical Face Masks**

Brand Name : A&Z MED

Model : LI - 2022

Type : II

and tested according to the following initial type tests by the manufacturer  
Technical standard EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods

For the assessment of conformity, the following documents were also supplied

Results of laboratory tests Ekoteks Testing Laboratory

Results of laboratory tests Ekoteks Testing Laboratory

Results of laboratory tests Ekoteks Testing Laboratory

UNIVERSAL CERTIFICATION has evaluated production, design, intended use, risk evaluation according to safety purpose, product itself and additional components (if exists) and product technical drawings of the medical face masks manufactured and designed for use during the medical operations or similar medical situations with same requirements which require restriction of infectious materials to be spread to patients. With this certificate, it is announced that the product fulfills all essential requirements and the related rules of 93/42/EEC Medical Devices Directive (MDD) Class I and II. The information on the packaging for the above listed products covers the necessary information stated in Annex I, §13, of the Medical Devices Directive (93/42/EEC) and Annex I, §21, of the Medical Device Regulation (EU) 2017/745. This information includes; reference to EN 14683 standard, type of mask (as indicated in Table 1) and other relevant information given in EN ISO 5223-1:2016 and EN 1041:2008+A1:2013. It is considered to be suitable to attach a CE mark, as seen below, on your products in accordance with the information given in this certificate with publishing an EU Declaration of Conformity.

This certificate is issued on 04/05/2020 and valid until 03/05/2021 with the conditions that no change has been made with the product references and no change in the production process or not suspended or withdrawn for any reason.

İSTANBUL -04/05/2020



Verify the validity with the QR Code



Suat KAÇMAZ  
UNIVERSAL CERTIFICATION  
Genel Müdür



# EU DECLARATION OF CONFORMITY

## PRODUCER

İBİŞLER TEKSTİL SANAYİ VE DİŞ TİCARET ANONİM ŞİRKETİ  
Orhangazi Mahallesi Tunç Caddesi No:5 34358 Esenyurt / İSTANBUL

## PRODUCT DESCRIPTION

Layered and molded medical device classified in the class I - Medical Device to be used as protection against inhalation of viruses, bacteria, other microorganisms, allergens from the environment (Type I and/or Type II)

The Producer / the Manufacturer declares his sole responsibility that the product above is, under conditions of normal use and conditions defined by the Producer / the Manufacturer, safe and meets all the necessary legal conditions and requirements. The product, a medical device that is intended for single use and solely in accordance with the Producer / the Manufacturer's instructions.

The Conformity is assessed especially with the following provisions:

- Government Regulation No. 93/42/EEC Medical devices establishing technical requirements for medical devices, in effect, wording
- Technical Standard EN 14683:2019+AC:2019 Medical face masks Requirements and test methods
- Other relevant harmonized legislation
- Other relevant international and community standards
- For the assessment of conformity, the following documents were also applied
- Tests for irritation and delayed-type hypersensitivity
- Results of laboratory tests Ekoteks Testing Laboratory BFE
- Results of laboratory tests Ekoteks Testing Laboratory Differential Pressure
- Results of laboratory tests Ekoteks Testing Laboratory Microbial Cleanliness

## MARKING, LABELING

Annex I, §13 of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745 specifies the information that should be specified on the packaging in which the medical face mask is supplied. The following information shall be supplied:

Type of mask (as indicated in Table 1). EN ISO 15223-1:2016 and EN 1041:2008+A1:2013 should be considered.

## MEASURES TO ENSURE CONFORMITY

The Producer / the Manufacturer declares that he has taken all necessary measures to ensure the conformity of products placed on the market with technical documentation and basic requirements for this type of product.

İlhan İBİŞ  
General Manager  
ISTANBUL 04/05/2020







## Certificate of Registration 2020

*This is to certify that the registration of*

**IBISLER TEKSTIL SANAYI VE DIS TICARET ANONIM SIRKETI**

**B BLOK, NO: 5 ORHAN GAZI MAHALLESİ TUNÇ CADDESİ, ESENYURT,**

**İSTANBUL, TÜRKİYE - 34522**

*with U.S. Food and Drug Administration as required by 21 CFR Part 80, is successfully completed by Liberty Management Group Ltd with the information provided by Ibisler Tekstil Sanayi Ve Dis Ticaret Anonim Sirketi*

<b>Owner/Operator Number</b>	<b>1007588</b>
<b>Date of Registration</b>	<b>May 22, 2020</b>
<b>Date of Expiration</b>	<b>December 31, 2020</b>
<b>US Agent</b>	<b>Liberty Management Group Ltd.</b>
<b>Device Listing Numbers</b>	<b>See Annex</b>
<b>Certificate Number</b>	<b>3005220120</b>

*This certificate does not make representations or warranties to any person or entity other than the named certificate holder; it is issued for record keeping purpose only. This certificate does not denote endorsement or approval of certificate holder's facility or product by the U.S. food and Drug Administration. Liberty management Group Ltd. assumes no liability to any person or entity in connection with the foregoing.*

*The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration. Liberty Management Group Ltd. is not affiliated with the U.S. Food and Drug Administration.*

**LMG** LIBERTY  
MANAGEMENT  
GROUP LTD.

75 Executive Drive, Aurora, Illinois, USA  
[www.fdahelp.us](http://www.fdahelp.us)

Manoj Zacharias

President

Liberty Management Group LTD.

Dated: June 11, 2020



## Certificate of Registration 2020

### Annex - Device Listings

Listing Number	Code	Device Name - Proprietary Names
D404264	OE/	Non-Surgical Isolation Gown - A&Z Med Gown A&Z Med Overall
D404263	QKR	Face mask (except N95 respirator for general public health care personnel per IIE guidance - A&Z Med Mask

This is to Certify that



A&Z A&ZMED

İBİŞLER TEKSTİL SAN. VE DİŞ TİC. A.Ş.

ORHANGAZİ MAH. TUNÇ CAD. NO:5/B ESENYURT / İSTANBUL, TURKEY

Conforms to the Requirements of

ISO 9001:2015

Quality Management System

Tulum ve Medikal Maske Dikimi ve Satışı.

*Jumpsuit and Medical Mask Sewing and Sale .*

Certificate Number : Q.020.080.TR

Expiry Date : 17.04.2023

Certification Period : 3 Years / 16.04.2023

Certified Date : 17.04.2020

Approving Officer:

*A. Cetgin*



HMI Certification Training Ltd.  
492 Bearwood Rd, Smethwick B66 4HB, Birmingham / West Midland - England  
This certificate remains the property of HMI it is validity is subject to arrangement  
between the certificated client and HMI. For further details go to  
[www.bvcsert.com](http://www.bvcsert.com)



This is to Certify that



A&Z A&ZMED

İBİŞLER TEKSTİL SAN. VE DIŞ TİC. A.Ş.

ORHANGAZİ MAH. TUNÇ CAD. NO:5/B ESENYURT / İSTANBUL, TURKEY

Conforms to the Requirements of

ISO 13485:2016

Medical Device Quality Management System

Tulum , Medikal Önlük ve Medikal Maske Dikimi ve Satışı.

*Jumpsuit, Medical Gowns and Medical Mask Sewing and Sale .*

Certificate Number : M.020.080.TR  
Certification Period : 3 Years / 16.04.2023


Expiry Date : 17.04.2023  
Certified Date : 17.04.2020

Approving Officer:

*A. Ötügen*







**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**  
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar  
İstanbul/ TÜRKİYE

**TEST REPORT**  
*DENEY RAPORU*

20013151-  
ing

04-20

**Customer name:** İBİŞLER TEKSTİL SAN. VE DİŞ TİC A.Ş.  
**Address:** ORHANGAZİ MAH. TUNÇ CAD. NO:5 HADIMKÖY  
**Buyer name:** -  
**Contact Person:** SALİH KAŞMER  
**Order No:** -  
**Article No:** -  
**Name and identity of test item:** Blue mask  
**The date of receipt of test item:** 16.04.2020  
**Re-submitted/re-confirmation date:** -  
**Date of test:** 16.04.2020-27.04.2020  
**Remarks:** -  
**Sampling:** The results given in this report belong to the received sample by vendor.  
**End-Use:** Men's & Women's&Kid's Wear / Mask  
**Care Label:** Not Specified  
**Number of pages of the report:** 5



**Date**  
27.04.2020

**Customer Representative**  
Hatice ACARALP

**Head of Testing Laboratory**  
Sevim A. RAZAK

27.04.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.  
Testing reports without signature and seal are not valid.




**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

20013151-  
ing

04-20

REQUIRED TESTS	RESULT	COMMENTS
<b>PHYSICAL PROPERTIES TESTS</b>		
Breathability (Differential Pressure) <sup>(3)</sup>	P	
<b>MICROBIOLOGICAL TEST</b>		
Bacterial Filtration Efficiency (BFE) <sup>(1)</sup>	P	Type 2
Microbial Cleanliness (Bioburden) <sup>(2)</sup>	P	
P: Pass F: Fail R: Refer to retailer technologist. <sup>(1)</sup> Test results were evaluated according to EN 14683:2019+AC:2019 Annex-B/Table- 1)limit values <sup>(2)</sup> Test results were evaluated according to EN ISO 11737-1:2018 limit values <sup>(3)</sup> Test results were evaluated according to EN 14683:2019+AC:2019 Annex -C/Table- 1 limit values		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95 %. Tests marked (\*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.  
Testing reports without signature and seal are not valid.



**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

20013151-  
ing

04-20

## TEST RESULTS

### BREATHABILITY (Differential Pressure)

**Test Method :** EKOTEKS 70 - Ref: EN 14683:2019+AC:2019 Annex-C (\*)

Test area is 25 mm in diameter , 5 different sample was taken

Adjusted airflow is 8 l/min. The differential pressure is read directly using a differential pressure manometer .

	<u>RESULT</u>	<u>REQUIREMENT</u>
<b>Differential Pressure)</b> (Pa/cm <sup>2</sup> )	24.6 Pa/cm <sup>2</sup>	< 40 Pa/cm <sup>2</sup> Type I and Type II mask

Gen.fl 36-2/03

**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

20013151-  
ing

04-20

## TEST RESULTS

### BACTERIAL FILTRATION EFFICIENCY (BFE)

Test Metod: EKOTEKS 70 (In-House Method-Bacterial Filtration Efficiency Testing –BFE /Ref: EN 14683:2019+AC:2019 Annex B Medical Face Masks,Requirements and Test Methods (\*))

A specimen of the mask material is clamped between a impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Test Flow Rate	28,3 L/dk
Test Flow Time	2 dakika
Sample Sizes	20x20 cm <sup>2</sup>
Microorganism	<i>Staphylococcus aureus</i> ATCC 6538
Bacterial concentration (cfu/ ml )	5x10 <sup>5</sup> kob/ ml
incubation conditions	24 saat, 35°C ± 2°C
Positive control sample average of number of Bacteria (C)	1.96x10 <sup>3</sup> kob/ ml

Gen.fl36-2/03

RESULTS			Requirement BFE (%)
Number of Test Sample	Test Sample (T) Number of Bacteria (cfu/ml)	Bacterial Filtration Efficiency ( % B )	
1	25	98.7 %	Type I ≥95  Type II ≥98
2	24	98.8 %	
3	24	98.8 %	
4	23	98.8 %	
5	23	98.8 %	

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen



**EKOTEKS LABORATUVAR ve GÖZETİM  
HİZMETLERİ A.Ş.**

20013151-  
ing

04-20

## TEST RESULTS

### MICROBIAL CLEANLINESS (Bioburden)

**Test Metod:** Ref: EN ISO 11737-1:2018 (\*)

The sample is put in extraciton liquid after shaking well, inoculated on the agar.  
After incubation at  $30 \pm 1$  ° C for 72 hours, growth microorganisms are counted on the agar.

	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/g)	24 cfu/g	$\leq 30$ cfu/g Type I and Type II mask

\*cfu= Colony forming unit.