


Dokumentation TYP IIR 3A Medical

TYP IIR 3A Medical

Klassenkennzeichnung	3 – lagige MNS TYP IIR
Typenkennzeichnung (Modell)	REF 2001
Zertifizierung	EN 14683:2019 TYP IIR Konformitätskennzeichen entsprechend der Regulation (MDD 93/42
Hersteller	3A Medical Products Co. Ltd. Yu An Industrial Park, 230001 Liu An PEOPLE'S REPUBLIC OF CHINA
Zugelassener Vertreter der EU	Shanghai International Holding Corp. GmbH Eiffestrasse 80, 20537 Hamburg
Haltbarkeit	2 Jahre bei Lagerung in feuchtigkeitsbeständigem, regenfreiem und belüftetem Raum.
Verpackung	50 Masken im Beutel in einer Box 40 Boxen im Karton = 2.000 Masken / Karton
Maße	Box (19cm*10cm*8cm) Karton (52cm*39,5cm*34cm)
Produktspezifikationen	3-Lagig Ohrschlaufen Latexfrei Material: SPP / Melt Blown Farbe: weiß / blau Einmalgebrauch
Produktabmessungen	17,5cm*9,5cm





3A MEDICAL

MEDICAL FACE MASK WITH EAR LOOPS

Latex-free

CE



EN14683 : 2019 TYPE IIR

REF 2001



QTY 50

3A MEDICAL


MEDICAL FACE MASK WITH EAR LOOPS

Latex-free



EN14683 : 2019 TYPE IIR

REF 2001



QTY 50

3A MEDICAL

MEDICAL FACE MASK WITH EAR LOOPS

REF 2001

LOT 200803



08.2022



3A Medical Products Co.,Ltd.
Yu An Industrial Park,230001 Liu An
PEOPLE'S REPUBLIC OF CHINA



(01)6971432939975(17)220831(10)200803

EC REP

European Representative:
Name: Shanghai International Holding Corp. GmbH(Europe)
Address: Eiffestrasse 80, D-20537, Hamburg, Germany
Linkman: JINLIANG
Tel: 0049-40-2513175 Fax: 0049-40-255726

Use Instructions:



- (GB) No mask completely removes the risk of contamination from blood or body fluids
- (DE) Keine Maske beseitigt das Risiko einer Kontamination durch Blut oder Körperflüssigkeiten vollständig
- (FR) Aucun masque ne supprime complètement le risque de contamination par le sang ou les fluides corporels
- (ES) Ninguna máscara elimina por completo el riesgo de contaminación de sangre o fluidos corporales
- (IT) Nessuna maschera rimuove completamente il rischio di contaminazione da sangue o fluidi corporei

Product Name: Medical Face Mask With Ear Loops

Product Size: 175*95mm

Valid Period: 2 Years

Package Quantity: 50 PCS

Storage Method: Please store in dry and ventilated environment between -20℃ and 40℃

Standard: EN14683:2019 Type IIR

Manufacturer: 3A Medical Products Co., Ltd.

Address: Yu An Industrial Park, 230001 Liu An. PEOPLE'S REPUBLIC OF CHINA

Contact No: 0564-3611700



Keep it dry



Avoid the sunlight



Please do not use it if package is damaged



Disposable

CE



Konformitätserklärung des Herstellers



安徽富美医疗科技有限公司

3A Medical Products Co., Ltd.

地址：安徽省六安市裕安区

Add: Yu An Industrial Park, Liu An City, P.R. China

Declaration of Conformity

Manufacturer: 3A Medical Products Co., Ltd

Address: Yu An Industrial Park, 237100, Liu An, PEOPLE'S REPUBLIC OF CHINA

European Representative:

Name: Shanghai International Holding Corp. GmbH(Hamburg)

Address: Eiffestrasse 80, 20537, Hamburg, Germany

Product name: Medical face mask

Classification: Class 1 Rule 1 of Annex IX of MDD 93/42/EEC)

Type: Type IIR

Rule: According to Rule I Annex VII

We here with declare that above mentioned products meet the requirement of the (MDD 93/42/EEC) Medical device directive and the following harmonized standards:

EN14683:2019

ENISO 15223-1:2016

ENISO10993-1:2009/AC: 2010

ENISO10993-10:2013

ENISO 14971:2012

ENISO1041:2008A1:2013

ENISO10993-5:2009

The Medical face mask meets the requirement of EN14683:2019

Signature

Name: Billy Zhang

Position: Management Representative

Place: Liu An City

Date: 2020-04-17



Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016 & EN ISO 13485:2016

This is to certify that: 3A Medical Products Co., Ltd
Yu An Industrial Park
Yu An district
Liu An
Anhui
237100
China

安徽富美医疗科技有限公司
中国
安徽省
六安市
裕安区
城南工业园
邮编: 237100

Holds Certificate No: **MD 731336**

and operates a Quality Management System which complies with the requirements of ISO 13485:2016 & EN ISO 13485:2016 for the following scope:

Manufacture and Distribution of following medical devices: sterile Surgical Mask, sterile Medical Mask, sterile Surgical Gown, sterile Medical Coverall, sterile Surgical Drape, sterile Surgical Pack, and non-sterile Isolation Gown, non-sterile Shoe Cover, non-sterile Pillow Cover, non-sterile Sleeve Cover, non-sterile Non-woven Swabs, non-sterile Surgical Cap, non-sterile Head Cover, non-sterile Fluorocover, non-sterile Scrub suit, non-sterile Bed Cover.

以下产品的生产和销售: 无菌的医用外科口罩、一次性使用医用口罩、一次性使用手术衣、医用一次性防护服、一次性使用手术铺单、一次性使用无菌手术包; 非无菌的隔离衣、医用隔离鞋套鞋套、枕套、袖套、无纺布纱布、卫生帽手术帽、医用帽、机头罩、消毒服、床罩。



For and on behalf of BSI:

Gary E Slack, Senior Vice President - Medical Devices

Original Registration Date: 2020-09-21

Latest Revision Date: 2020-09-21

Effective Date: 2020-09-21

Expiry Date: 2023-09-20

Page: 1 of 1



...making excellence a habit.™

DIMDI Eintragung

Anlage 1
(zu § 4 Abs. 1 Nr. 1 DIMDIV)
Formularnummer 00304117

Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika
Form for Medical Devices except In Vitro Diagnostic Medical Devices

Zuständige Behörde / Competent authority		
Code	DE/CA05	
Bezeichnung / Name	Behörde für Gesundheit und Verbraucherschutz, Referat V43	
Staat / State	Land / Federal state	Deutschland Hamburg
Ort / City	Postleitzahl / Postal code	Hamburg 20539
Straße, Haus-Nr. / Street, house no. Billstraße 80		
Telefon / Phone	Telefax / Fax	+49-40-428280 +49-40-427310017
E-Mail / E-mail medizinprodukte@bgv.hamburg.de		

Anzeige / Notification		
Registrierdatum bei der zuständigen Behörde Registration date at competent authority	Registriernummer / Registration number	
Typ der Anzeige / Notification type S Erstanzeige / Initial notification E Änderungsanzeige / Notification of change E Widerrufsanzeige / Notification of withdrawal		
Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn		
Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG E Hersteller / Manufacturer S Bevollmächtigter / Authorised Representative E Einführer / Importer E Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG E Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV E Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG		

Anzeigender / Reporting organisation (person)		
Code	DE/0000040627	
Bezeichnung / Name	Shanghai International Holding Corporation GmbH (Europe)	
Staat / State	Land / Federal state	
Deutschland	Hamburg	
Ort / City	Postleitzahl / Postal code	
Hamburg	20537	
Straße, Haus-Nr. / Street, house no.		
Eiffestrasse 80		
Telefon / Phone	Telefax / Fax	
+49-40-2513175	+49-40-255726	
E-Mail / E-mail		
shholding@hotmail.com		

Hersteller / Manufacturer		
Bezeichnung / Name	3A Medical Products Co., Ltd.	
Staat / State	CN	
Ort / City	Postleitzahl / Postal code	
Liu An,	230001	
Straße, Haus-Nr. / Street, house no.		
Yu An Industrial Park,		
Telefon / Phone	Telefax / Fax	
+86-564-3611700	+86-564-3611700	
E-Mail / E-mail		
guweilin@3a-medical.cn		

Sicherheitsbeauftragter für Medizinprodukte nach § 30 Abs. 2 MPG 9) Safety officer for medical devices pursuant to § 30 (2) Medical Devices Act, MPG		
Bezeichnung / Name	Liang Jin	
Staat / State	Land / Federal state	
Deutschland	Hamburg	
Ort / City	Postleitzahl / Postal code	
Hamburg	20537	
Straße, Haus-Nr. / Street, house no.		
Eiffestr.80		
Telefon / Phone	Telefax / Fax	
+49-40-2513175	+49-40-255726	
E-Mail / E-mail		
shholding@hotmail.com		

Vertreter / Deputy (optional)		
	Bezeichnung / Name	
	Telefon / Phone	Telefax / Fax
	E-Mail / E-mail	
	E Erstanzeige / Initial notification S Änderungsanzeige / Notification of change	

Medizinprodukt (Erstmaliges Inverkehrbringen) / Medical device (First placing on the market)	
Klasse / Class	
S I	
E I - steril / sterile	
E I - mit Messfunktion / with measuring function	
E I - steril und mit Messfunktion / sterile and with measuring function	
E IIa	
E IIb	
E III	
E III - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012	
manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012	
E Aktives implantierbares Medizinprodukt / Active implantable medical device	
E Aktives implantierbares Medizinprodukt - hergestellt unter Verwendung von Gewebe tierischen Ursprungs im Sinne der Verordnung (EU) Nr. 722/2012	
Active implantable medical device - manufactured utilising tissues of animal origin in terms of Commission Regulation (EU) No 722/2012	
App (Software auf mobilen Endgeräten)	E ja / yes S nein / no
Nummer(n) der Bescheinigung(en) / Certificate number(s)	
Handelsname des Produktes / Trade name of the device	
3A	
Produktbezeichnung / Name of device	
Medical face mask	
Nomenklaturcode / Nomenclature code	
12-447	
Nomenklaturbezeichnung / Nomenclature term	
Maske	
Kategoriecode / Category code	
10	
Kategorie / Category	
Produkte zum Einmalgebrauch	
Kurzbeschreibung deutsch / German short description	
Kurzbeschreibung englisch / English short description	
Medical face mask is intend to be used for hygiene care which the wearer under general medical environment where there is no risk of bodily fluids splash.	

Medizinprodukte (Aufbereiten) / Medical devices (Reprocessing)	
	E. Semikritische Medizinprodukte / Semicritical medical devices E. Gruppe A / Group A E. Gruppe B / Group B
	E. Kritische Medizinprodukte / Critical medical devices E. Gruppe A / Group A E. Gruppe B / Group B E. Gruppe C / Group C Nummer der Bescheinigung / Certificate number
	Sterilisationsverfahren / Sterilisation procedures E. Dampfsterilisation / Steam sterilisation E. Gassterilisation / Gas sterilisation E. Strahlensterilisation / Radiation sterilisation E. andere / others Angewandtes Verfahren / Applied procedure

Ich versichere, dass die Angaben nach bestem Wissen und Gewissen gemacht wurden.
I affirm that the information given above is correct to the best of my knowledge.

Ort City Hamburg	Datum Date 2020-04-30
	Name Liang Jin
Unterschrift Signature	

Bearbeitungsvermerke / Processing notes	
Nur von der zuständigen Behörde auszufüllen / To be filled in only by the competent authority	
Bearbeiter / Person responsible	Telefon / Phone

CERTIFICATE OF NOTIFICATION

This is to certify that, according to European Council Directive 93/42/EEC, Shanghai International Holding Corp, GmbH (Europe), performed all notification duties and responsibilities as the European authorized Representative:

MANUFACTURER: 3A Medical Products Pty Ltd

Address: Yu An Industrial Park, 230001, Liu An, CN

The manufacturer has provided Shanghai International Holding Corp, GmbH (Europe), with all the appropriate declaration according to the European Council Directive 93/42 EEC including the EC Declaration of Conformity confirming that its medical device, as stipulated here below, is fulfilling the essential requirements of the European Council Directive 93/42/EEC.

Devices: Medical Face Mask


Classification: I

Model: 17.5(±5%)cm*9.5(±5%)cm, 14.5(±5%)cm*9.5(±5%)cm

Where the manufacturer affix the CE mark to the device listed they must ensure that all the essential requirements of European Council Directive 93/42/EEC are met.

The notification of abovementioned device has been completed by the European Authorized Representative in Germany. The Germany Competent Authority is notified of the manufacture's device and has allocated registration.

EXECUTIVE
DIRECTOR



Shine

Test Report



**3A Medical Products Co. Ltd, Yu An
Industrial Park, 230001 Liu An. PR
China**

Your notice of
17-03-2020

Your reference

Date
03-04-2020

Analysis Report 20.01605.03

Required tests :

EN 14683 (2019) + AC (2019)	EN 14683 - annex B (2019) + AC (2019)	Bacterial filtration efficiency
EN 14683 (2019) + AC (2019)	ISO 22609 (2004)	Medical face masks - Splash Test
EN 14683 (2019) + AC (2019)	EN 14683 - annex C (2019) + AC (2019)	Medical face masks - Breathability (differential pressure)
EN 14683 (2019) + AC (2019)	EN 14683 - §5.2.5 (2019) AC (2019)	Microbial cleanliness on masks

Identification number	Information given by the client	Date of receipt
T2006058	REF 2001 Lot 202001202	17-03-2020

Sylvie Niessen
Order responsible

This report may be reproduced, as long as it is presented in its entire form, without written permission of Centexbel.
The results of the analysis cover the received samples. Centexbel is not responsible for the representativeness of the samples.
In assessing compliance with the specifications, we did not take into account the uncertainty on the test results.

INRICHTING ERKEND BIJ TOEPASSING VAN DE BESLUITWET VAN 30 JANUARI 1947 / ETABLISSEMENT RECONNU PAR APPLICATION DE L'ARRÊTÉ-LOI DU 30 JANVIER 1947



CENTEXBEL • textile competence centre • www.centexbel.be • www.vkc.be

GENT • Technologiepark 70 • BE-9052 Zwijnaarde, Belgium • phone +32 9 220 41 51 • fax +32 9 220 49 55 • gent@centexbel.be
GRÂCE-HOLLOGNE • Rue du Travail 5 • BE-4460 Grâce-Hollogne, Belgium • phone +32 4 296 82 00 • g-h@centexbel.be
KORTRIJK • Etienne Sabbelaan 49 • BE-8500 Kortrijk, Belgium • phone +32 56 29 27 00 • fax +32 56 29 27 01 • info@vkc.be
VAT BE 0459.218.289 • IBAN BE44 2100 4729 6545 • BIC GEBABEBB



Reference: T2006058 - REF 2001
Lot 202001202

Bacterial filtration efficiency

Date of ending the test	25-03-2020
Standard used	EN 14683 - annex B (2019) + AC (2019)
Product standard	EN 14683 (2019) + AC (2019)
Mask description	3 layers/ 18gsm SPP white/ 25gsm/meltex/ 22gsm SPP blue
Number of tested masks :	5
BFE Area tested :	$\pm 49 \text{ cm}^2$
Masks conditioning :	$21 \pm 5^\circ\text{C}$ and $85 \pm 5\% \text{ RH}$
Side of the mask in contact with the bacterial challenge :	Inner side
Challenge bacterial strain used :	<i>Staphylococcus aureus</i> ATCC6538
Bacterial challenge per test :	1700 - 3000 CFU
Total test time :	1 min. delivering challenge + 1 min. without challenge (air flow continuing)
Flow rate :	28.3 l/min.
Positive control	Tests performed with no filter material in the air stream
Negative control	Test performed without challenge



Results

B = Bacterial filtration efficiency (%)

$$B = \frac{(C - T)}{C} \times 100$$

With C = mean of the total plate counts for the positive control runs
T = total count for the tested mask

# Mask	B (%)
1	99.9
2	99.2
3	99.8
4	99.9
5	99.8

Mean particle size of the bacterial challenge aerosol : 2.9 µm

Controls

Mean positive controls 2636 CFU
Negative control < 1 CFU



Reference: T2006058 - REF 2001
 Lot 202001202

Medical face masks - Splash Test

Date of ending the test	26-03-2020
Standard used	ISO 22609 (2004)
Product standard	EN 14683 (2019) + AC (2019)
Mask description	3 layers/ 18gsm SPP white/ 25gsm/meltex/ 22gsm SPP blue
Number of tested masks :	32
Blood surface tension	42 ± 2 dynes/cm
Volume of the delivered blood	2 ml
Distance "canula-mask"	30 ± 1 cm
Side of the mask "impacted"	Outer side
Masks conditioning :	21 ± 5°C and 85 ± 5% RH

Results

Blood pressure tested 16.0 kPa

Controls

Blood visualisation on the mask	OK
Calibration procedure	OK
Control of the blood volume delivered (2 ml)	
- before the test :	OK
- after 16 masks :	OK
- after 32 masks :	OK



Results obtained on the set of masks

# Mask	Results : pass / fail
1	Pass
2	Pass
3	Pass
4	Pass
5	Pass
6	Pass
7	Pass
8	Pass
9	Pass
10	Pass
11	Pass
12	Pass
13	Pass
14	Pass
15	Pass
16	Pass
17	Pass
18	Pass
19	Pass
20	Pass
21	Pass
22	Pass
23	Pass
24	Pass
25	Pass
26	Pass
27	Pass
28	Pass
29	Pass
30	Pass
31	Pass
32	Pass



Summary P = 16.0 kPa

Number of "Pass" masks	Number of "Fail" masks
32	0

Pass = no blood detected on the observed side

Fail = blood detected on the observed side

In agreement with the customer the number of tested mask has been determined based on a single sampling plan providing an AQL of 4 % (acceptable quality limit).


If 29 masks or more over 32 obtain a "Pass" result the 4% AQL is reached.

Reference: T2006058 - REF 2001
 Lot 202001202

Medical face masks - Breathability (differential pressure)

Date of ending the test	24-03-2020
Standard used	EN 14683 - annex C (2019) + AC (2019)
Product standard	EN 14683 (2019) + AC (2019)
Mask description	3 layers/ 18gsm SPP white/ 25gsm/meltex/ 22gsm SPP blue
Number of tested masks :	5
Number of areas per mask	5 (see figure)
Dimension of the areas :	Disc whose diameter is 2.5 cm
Surface areas :	4.9 cm ²
Flow rate :	8 l/min.
Direction of the air flow :	From the inside of the mask to the outside
Masks conditioning :	21 ± 5°C and 85 ± 5% RH

Figure : Distribution of the areas in the mask





Results

ΔP

	Mask 1	Mask 2	Mask 3	Mask 4	Mask 5
Area A	52.6	23.0	52.6	35.2	40.3
Area B	51.1	30.6	59.1	39.7	39.7
Area C	53.0	34.8	45.4	43.4	38.3
Area D	46.7	35.7	48.1	44.2	41.4
Area E	44.2	29.7	39.5	39.7	41.4
Average ΔP (Pa/cm²)	49.5	30.8	48.9	40.4	40.2



Reference: T2006058 - REF 2001
 Lot 202001202

Microbial cleanliness on masks

Date of ending the test 30-03-2020
 Standard used EN 14683 - §5.2.5 (2019) AC (2019)
 Product standard EN 14683 (2019) + AC (2019)

 Number of tested masks 5
 Extraction liquid Peptone 1g/l, NaCl 5g/l & Tween 20 2g/l
 Extraction volume 300 ml
 Extraction time 5 min.
 Counting technique Membrane filtration
 Filtration volume 100 ml
 Culture media TSA (Tryptic Soy Agar)
 SDA (Sabouraud Dextrose Agar with chloramphenicol)
 Incubation conditions 3 days at 30°C (TSA)
 7 days at 20-25°C (SDA)

Results

# Mask	Mask weight (g)	CFU*/mask		Microbial cleanliness	
		<i>Aerobic microbial count (bacteria)</i>	<i>Fungi count (SDA)</i>	Σ CFU/mask	Σ CFU/g
1	4.25	33	<3	< 36	< 9
2	4.22	3	<3	< 6	< 2
3	4.33	3	<3	< 6	< 2
4	4.35	36	<3	< 39	< 9
5	4.37	3	<3	< 6	< 2